Environment Information Technology Pty Ltd
ABN 23 116 474 634

Unit 1, 19 Kays Lane
Alstonville NSW 2477
Phone (02) 66283400
www.eitechnology.com.au

Products affected	IPC Series 2 (Atmel)
Serial numbers	All
Action required	None – Advisory only
Reason for notice	Users wishing to change the operation of an IPC from SDI-12 to RS485 and vise-versa.
Notes	Conversion of Series 1 IPC (PIC) is not available as a field upgrade. These units must be returned to the factory for upgrade. Series 1 are identified by having two green terminal block connectors on the front panel. Series 2 have one green terminal block and two smaller blue connectors for DTN and EXPN.
Resolution	To change the mode of operation of an IPC it is necessary to change the setting of header pin jumpers on the IPC circuit board. This requires removing the four corner front panel screws, and setting the JP5 jumpers as shown below:

SDI-12

::II:

RS485

II:::



Ensure that front panel screws are firmly in place before returning to service.